

# Ican Autoclave Certificate of Validation



Registered business	Job address	Landlord/agent
Excel Health Care Limited Unit 5 Childerditch Industrial Park Childerditch Hall Drive Little Warley Brentwood Essex CM13 3HD +44 0208 597 1404	18 BAR STREET SCARBOROUGH NORTH YORKSHIRE YO11 2HT	N/A
Question		
Certificate of validation, a series of tests carried out as instructed under HTM 01-05 guidelines. This certificate is valid for 1 year only from the date of testing.		
Autoclave Type & Model		
Enigma 8LB		
Serial Number		
32197		
Max W.P.		
2.8 BAR		
Automatic Air Leakage Test / Certified fit for (E13)		
Passed		
Residual Air Test / Certified fit for use (E5)		
Passed		
Automatic Control Test / Certified fit for use (E4)		
Passed		
Steam Penetration Test / Advised customer in the use of		
Passed		
Thermometric Test / Certified fit for use (E1/E2/E3)		
Passed		
Air Leakage Test / Advised customer if fitted		
Passed		
Steam Generator cut out test / Certified fit for use		


Ican Autoclave Certificate of Validation (28389)

Registered business :Excel Health Care Limited

Registered business address :Unit 5 Childerditch Industrial Park Childerditch Hall Drive Little Warley Brentwood Essex CM13 3HD

# Ican Autoclave Certificate of Validation



Question		
Passed		
Test Date		
16/04/26		
Engineer Name		
Cole Hopton		
Engineer information	Received by information	Certificate information
Engineer name: Cole Hopton (DN11) Signature: 	Name: HEARTS & CRAFTS Received by: Customer Signature: 	Date: 16/04/2026 Number : 28389